

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070419\
 Data File : VN056602.D
 Acq On : 3 Jul 2019 22:07
 Operator : JC/SP
 Sample : K3680-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2R

Quant Time: Jul 03 23:58:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	274376	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	478756	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	484518	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	189102	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	160572	52.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.60%	
35) Dibromofluoromethane	7.59	113	140069	46.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.80%	
50) Toluene-d8	10.09	98	600801	51.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.40%	
62) 4-Bromofluorobenzene	12.40	95	213822	55.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.38%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.80	59	262800	513.110	ug/l	98
16) Acetone	3.82	43	4026	2.643	ug/l	98
17) Carbon Disulfide	4.05	76	3183	0.577	ug/l #	92
19) Methyl tert-butyl Ether	5.05	73	196511	24.502	ug/l	99
85) sec-Butylbenzene	13.17	105	10774	0.708	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070419\
Data File : VN056602.D
Acq On : 3 Jul 2019 22:07
Operator : JC/SP
Sample : K3680-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2R

Quant Time: Jul 03 23:58:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

