

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070518\
 Data File : VN049715.D
 Acq On : 5 Jul 2018 16:07
 Operator : MD\SY
 Sample : PB110828ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110828ZHE#02

Quant Time: Jul 05 16:50:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1248662	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2022446	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1659937	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	613830	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	850942	51.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.20%	
35) Dibromofluoromethane	7.59	113	702714	46.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.96%	
50) Toluene-d8	10.09	98	2682460	45.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.66%	
62) 4-Bromofluorobenzene	12.40	95	762397	37.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.46%	
Target Compounds						
16) Acetone	3.83	43	34684	2.565	ug/l	99
20) Methylene Chloride	4.55	84	20334	1.198	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070518\
Data File : VN049715.D
Acq On : 5 Jul 2018 16:07
Operator : MD\SY
Sample : PB110828ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110828ZHE#02

Quant Time: Jul 05 16:50:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

