

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070518\
 Data File : VN049716.D
 Acq On : 5 Jul 2018 16:33
 Operator : MD\SY
 Sample : PB110828ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110828ZHE#03

Quant Time: Jul 05 17:00:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1213799	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2020244	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1678876	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	623073	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	858668	53.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.14%	
35) Dibromofluoromethane	7.59	113	699839	46.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.68%	
50) Toluene-d8	10.09	98	2663324	45.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.10%	
62) 4-Bromofluorobenzene	12.41	95	785641	38.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.84%	
Target Compounds						
16) Acetone	3.83	43	36814	3.560	ug/l	96
20) Methylene Chloride	4.55	84	21760	1.318	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070518\
Data File : VN049716.D
Acq On : 5 Jul 2018 16:33
Operator : MD\SY
Sample : PB110828ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110828ZHE#03

Quant Time: Jul 05 17:00:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

