

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070518\
 Data File : VN049720.D
 Acq On : 5 Jul 2018 18:17
 Operator : MD\SY
 Sample : PB110828ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110828ZHE#07

Quant Time: Jul 06 02:18:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1179890	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1985502	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1609298	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	604017	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	839597	53.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.76%	
35) Dibromofluoromethane	7.59	113	684969	46.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.30%	
50) Toluene-d8	10.09	98	2604235	44.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.64%	
62) 4-Bromofluorobenzene	12.40	95	745050	37.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.12%	
Target Compounds						
16) Acetone	3.83	43	33688	2.867	ug/l	95
18) Methyl Acetate	4.33	43	94913	1.605	ug/l	98
20) Methylene Chloride	4.55	84	20362	1.269	ug/l	94
25) 2-Butanone	6.85	43	8418	1.627	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070518\
Data File : VN049720.D
Acq On : 5 Jul 2018 18:17
Operator : MD\SY
Sample : PB110828ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110828ZHE#07

Quant Time: Jul 06 02:18:12 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

