

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070522\
 Data File : VN073318.D
 Acq On : 05 Jul 2022 16:23
 Operator : JC\MD
 Sample : N3564-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14B-11-1

Quant Time: Jul 06 01:15:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	276896	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	477043	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	449222	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	228985	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	181138	45.315	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.640%
35) Dibromofluoromethane	7.951	113	151862	47.848	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.700%
50) Toluene-d8	10.380	98	512268	45.444	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.880%
62) 4-Bromofluorobenzene	12.668	95	229096	52.885	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.780%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	59952	26.720	ug/l	91
20) Methylene Chloride	5.016	84	206357	52.332	ug/l	89
31) Cyclohexane	8.016	56	13925	2.127	ug/l #	67
37) Ethyl Acetate	7.345	43	44662	6.053	ug/l	98
39) Methylcyclohexane	9.398	83	52532	8.640	ug/l	93
43) Isopropyl Acetate	8.492	43	201333	18.985	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070522\
Data File : VN073318.D
Acq On : 05 Jul 2022 16:23
Operator : JC\MD
Sample : N3564-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14B-11-1

Quant Time: Jul 06 01:15:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

