

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070522\
 Data File : VN073319.D
 Acq On : 05 Jul 2022 16:47
 Operator : JC\MD
 Sample : N3564-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14B-11-1-GR

Quant Time: Jul 06 01:58:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	272412	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	462591	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	431134	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	217902	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	172685	43.912	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.820%
35) Dibromofluoromethane	7.951	113	148784	48.343	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.680%
50) Toluene-d8	10.380	98	487704	44.617	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.240%
62) 4-Bromofluorobenzene	12.669	95	223044	53.097	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.200%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	66336	30.051	ug/l	90
20) Methylene Chloride	5.016	84	132986	34.281	ug/l	92
31) Cyclohexane	8.022	56	161315	25.047	ug/l	90
39) Methylcyclohexane	9.392	83	116414	19.745	ug/l	96
43) Isopropyl Acetate	8.492	43	211169	20.535	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070522\
Data File : VN073319.D
Acq On : 05 Jul 2022 16:47
Operator : JC\MD
Sample : N3564-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14B-11-1-GR

Quant Time: Jul 06 01:58:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

