

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070522\
 Data File : VN073321.D
 Acq On : 05 Jul 2022 17:36
 Operator : JC\MD
 Sample : N3564-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14B-11-2-GR

Quant Time: Jul 06 01:15:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	283731	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	479358	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	448087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	230745	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	179238	43.760	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.520%
35) Dibromofluoromethane	7.951	113	151355	47.458	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.920%
50) Toluene-d8	10.381	98	505672	44.643	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.280%
62) 4-Bromofluorobenzene	12.669	95	230806	53.023	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.040%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	67563	29.386	ug/l	94
20) Methylene Chloride	5.016	84	219975	54.442	ug/l	88
25) 2-Butanone	7.269	43	19424	5.306	ug/l	97
31) Cyclohexane	8.022	56	31817	4.743	ug/l #	86
37) Ethyl Acetate	7.345	43	54574	7.360	ug/l	97
39) Methylcyclohexane	9.392	83	42005	6.875	ug/l	91
43) Isopropyl Acetate	8.498	43	214043	20.086	ug/l #	86
73) Isopropylbenzene	12.516	105	33618	1.687	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070522\
Data File : VN073321.D
Acq On : 05 Jul 2022 17:36
Operator : JC\MD
Sample : N3564-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14B-11-2-GR

Quant Time: Jul 06 01:15:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

