

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070522\
 Data File : VN073322.D
 Acq On : 05 Jul 2022 18:01
 Operator : JC\MD
 Sample : N3564-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14B-11-3

Quant Time: Jul 06 01:16:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	282145	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	481765	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	435766	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	207381	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	177037	43.465	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.940%		
35) Dibromofluoromethane	7.951	113	151378	47.228	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.460%		
50) Toluene-d8	10.381	98	500559	43.971	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	87.940%		
62) 4-Bromofluorobenzene	12.669	95	219726	50.225	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	100.460%		
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	73273	32.049	ug/l	93
18) Methyl Acetate	4.793	43	20666	3.226	ug/l #	84
20) Methylene Chloride	5.016	84	256048	63.726	ug/l	92
31) Cyclohexane	8.022	56	48041	7.202	ug/l #	93
37) Ethyl Acetate	7.340	43	50336	6.755	ug/l	95
39) Methylcyclohexane	9.392	83	97170	15.825	ug/l	92
43) Isopropyl Acetate	8.498	43	203674	19.017	ug/l #	86
73) Isopropylbenzene	12.516	105	61236	3.419	ug/l	99
78) n-propylbenzene	12.857	91	52120	2.620	ug/l	98
85) sec-Butylbenzene	13.439	105	19457	1.107	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070522\
Data File : VN073322.D
Acq On : 05 Jul 2022 18:01
Operator : JC\MD
Sample : N3564-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14B-11-3

Quant Time: Jul 06 01:16:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

