

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070522\
 Data File : VN073323.D
 Acq On : 05 Jul 2022 18:25
 Operator : JC\MD
 Sample : PB146029TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146029TB

Quant Time: Jul 06 01:16:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	273239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	462942	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	431758	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	209421	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	173943	44.098	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.200%
35) Dibromofluoromethane	7.951	113	146574	47.589	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.180%
50) Toluene-d8	10.381	98	494115	45.169	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.340%
62) 4-Bromofluorobenzene	12.669	95	218336	51.937	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.880%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	36145	16.325	ug/l #	88
18) Methyl Acetate	4.787	43	7348	1.184	ug/l #	64
20) Methylene Chloride	5.022	84	33738	8.671	ug/l #	89
37) Ethyl Acetate	7.340	43	62316	8.702	ug/l	95
43) Isopropyl Acetate	8.492	43	245560	23.861	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070522\
Data File : VN073323.D
Acq On : 05 Jul 2022 18:25
Operator : JC\MD
Sample : PB146029TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146029TB

Quant Time: Jul 06 01:16:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

