

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070522\
 Data File : VN073326.D
 Acq On : 05 Jul 2022 19:39
 Operator : JC\MD
 Sample : N3564-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-16-2-6-GR

Quant Time: Jul 06 01:17:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	290471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	496702	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	460078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	229819	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	183966	43.872	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	87.740%		
35) Dibromofluoromethane	7.951	113	156608	47.390	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.780%		
50) Toluene-d8	10.381	98	525050	44.735	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	89.460%		
62) 4-Bromofluorobenzene	12.669	95	234985	52.098	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	104.200%		
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	48858	20.757	ug/l	93
20) Methylene Chloride	5.016	84	266453	64.415	ug/l	92
25) 2-Butanone	7.269	43	37872	10.106	ug/l #	89
37) Ethyl Acetate	7.345	43	46330	6.030	ug/l	96
43) Isopropyl Acetate	8.492	43	195321	17.689	ug/l #	87
52) Toluene	10.439	92	33721	3.477	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070522\
Data File : VN073326.D
Acq On : 05 Jul 2022 19:39
Operator : JC\MD
Sample : N3564-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-16-2-6-GR

Quant Time: Jul 06 01:17:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

