

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070522\
 Data File : VN073328.D
 Acq On : 05 Jul 2022 20:28
 Operator : JC\MD
 Sample : N3564-25
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-16-2-5-GR

Quant Time: Jul 06 01:17:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	277360	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	478559	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	449355	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	227650	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	177920	44.436	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.880%
35) Dibromofluoromethane	7.951	113	152775	47.983	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.960%
50) Toluene-d8	10.380	98	508650	44.981	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.960%
62) 4-Bromofluorobenzene	12.668	95	226766	52.182	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.360%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	55665	24.767	ug/l	94
18) Methyl Acetate	4.793	43	6637	1.054	ug/l #	81
20) Methylene Chloride	5.016	84	122266	30.955	ug/l	89
25) 2-Butanone	7.269	43	34343	9.598	ug/l	90
37) Ethyl Acetate	7.339	43	58525	7.906	ug/l	96
43) Isopropyl Acetate	8.492	43	222360	20.901	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070522\
Data File : VN073328.D
Acq On : 05 Jul 2022 20:28
Operator : JC\MD
Sample : N3564-25
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-16-2-5-GR

Quant Time: Jul 06 01:17:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

