

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070618\
 Data File : VN049734.D
 Acq On : 6 Jul 2018 13:11
 Operator : MD\SY
 Sample : J3840-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-6-D

Quant Time: Jul 06 14:34:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1173386	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1984533	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1649434	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	594865	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	855152	55.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.38%	
35) Dibromofluoromethane	7.59	113	705539	47.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.10%	
50) Toluene-d8	10.09	98	2669448	45.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.94%	
62) 4-Bromofluorobenzene	12.40	95	761148	38.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.78%	
Target Compounds						
16) Acetone	3.83	43	40936	5.335	ug/l	99
20) Methylene Chloride	4.55	84	166238	10.418	ug/l	96
95) Naphthalene	15.14	128	262499	14.028	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070618\
Data File : VN049734.D
Acq On : 6 Jul 2018 13:11
Operator : MD\SY
Sample : J3840-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-6-D

Quant Time: Jul 06 14:34:19 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

