

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070618\
 Data File : VN049740.D
 Acq On : 6 Jul 2018 15:46
 Operator : MD\SY
 Sample : J3887-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-8-S

Quant Time: Jul 06 17:13:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1187488	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1937732	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1629248	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	631261	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	828301	52.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.64%	
35) Dibromofluoromethane	7.59	113	689100	47.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.14%	
50) Toluene-d8	10.09	98	2624016	46.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.56%	
62) 4-Bromofluorobenzene	12.40	95	758856	39.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.40%	
Target Compounds						
16) Acetone	3.83	43	39999	4.866	ug/l	96
20) Methylene Chloride	4.55	84	127500	7.896	ug/l	97
25) 2-Butanone	6.85	43	12416	2.385	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070618\
Data File : VN049740.D
Acq On : 6 Jul 2018 15:46
Operator : MD\SY
Sample : J3887-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-8-S

Quant Time: Jul 06 17:13:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

