

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070618\
 Data File : VN049743.D
 Acq On : 6 Jul 2018 17:04
 Operator : MD\SY
 Sample : PB110875TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110875TB

Quant Time: Jul 06 18:38:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1215051	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1977977	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1655347	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	591498	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	840930	52.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.82%	
35) Dibromofluoromethane	7.59	113	648883	43.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.76%	
50) Toluene-d8	10.09	98	2654849	45.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.74%	
62) 4-Bromofluorobenzene	12.40	95	768711	38.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.80%	
Target Compounds						
16) Acetone	3.82	43	33155	2.374	ug/l	98
20) Methylene Chloride	4.55	84	32318	1.956	ug/l	95
95) Naphthalene	15.14	128	11244	4.470	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070618\
Data File : VN049743.D
Acq On : 6 Jul 2018 17:04
Operator : MD\SY
Sample : PB110875TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110875TB

Quant Time: Jul 06 18:38:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

