

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070618\
 Data File : VN049746.D
 Acq On : 6 Jul 2018 18:21
 Operator : MD\SY
 Sample : PB110875ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110875ZHE#03

Quant Time: Jul 06 22:59:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1177293	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1945996	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1601492	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	609918	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	810206	52.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.22%	
35) Dibromofluoromethane	7.59	113	649838	44.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.34%	
50) Toluene-d8	10.09	98	2558722	44.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.86%	
62) 4-Bromofluorobenzene	12.40	95	730650	37.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.16%	
Target Compounds						
16) Acetone	3.83	43	31723	2.242	ug/l	100
18) Methyl Acetate	4.33	43	89096	1.148	ug/l	99
20) Methylene Chloride	4.55	84	28529	1.782	ug/l	97
25) 2-Butanone	6.85	43	10902	2.112	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070618\
Data File : VN049746.D
Acq On : 6 Jul 2018 18:21
Operator : MD\SY
Sample : PB110875ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110875ZHE#03

Quant Time: Jul 06 22:59:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

