

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070618\
 Data File : VN049748.D
 Acq On : 6 Jul 2018 19:13
 Operator : MD\SY
 Sample : J3869-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-1

Quant Time: Jul 09 09:15:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1495242	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2523475	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2099172	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	746935	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	1056487	53.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.00%	
35) Dibromofluoromethane	7.59	113	875215	46.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.78%	
50) Toluene-d8	10.09	98	3366254	45.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.18%	
62) 4-Bromofluorobenzene	12.40	95	976260	38.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	77.44%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070618\
Data File : VN049748.D
Acq On : 6 Jul 2018 19:13
Operator : MD\SY
Sample : J3869-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-1

Quant Time: Jul 09 09:15:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

