

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070618\
 Data File : VN049750.D
 Acq On : 6 Jul 2018 20:05
 Operator : MD\SY
 Sample : J3869-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TW-3

Quant Time: Jul 09 09:16:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1521263	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2520030	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2108924	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	781230	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	1055353	52.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.06%	
35) Dibromofluoromethane	7.59	113	904541	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
50) Toluene-d8	10.09	98	3387521	45.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.88%	
62) 4-Bromofluorobenzene	12.40	95	988629	39.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.54%	
Target Compounds						
19) Methyl tert-butyl Ether	5.05	73	16630	0.361	ug/l	97
30) Chloroform	7.37	83	1176563	32.994	ug/l	99
47) Bromodichloromethane	9.40	83	75557	2.623	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070618\
Data File : VN049750.D
Acq On : 6 Jul 2018 20:05
Operator : MD\SY
Sample : J3869-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TW-3

Quant Time: Jul 09 09:16:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

