

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070620\
 Data File : VN062391.D
 Acq On : 7 Jul 2020 2:48
 Operator : JC/MD
 Sample : PB129462TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129462TB

Quant Time: Jul 07 05:24:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 07 05:08:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	905746	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	1705924	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	1672532	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	609702	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	822672	50.21	ug/l	0.00
Spiked Amount			Recovery	=	100.42%	
35) Dibromofluoromethane	7.55	113	566551	48.68	ug/l	0.00
Spiked Amount			Recovery	=	97.36%	
50) Toluene-d8	10.06	98	2185612	50.24	ug/l	0.00
Spiked Amount			Recovery	=	100.48%	
62) 4-Bromofluorobenzene	12.38	95	780488	46.84	ug/l	0.00
Spiked Amount			Recovery	=	93.68%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
8) Diethyl Ether	3.38	74	2200	0.317	ug/l	89
11) Tert butyl alcohol	4.75	59	1264	0.557	ug/l #	89
13) Acrolein	3.56	56	125	0.466	ug/l #	12
16) Acetone	3.79	43	61851	8.188	ug/l	94
18) Methyl Acetate	4.28	43	18967	0.729	ug/l	96
20) Methylene Chloride	4.50	84	149933	12.457	ug/l	96
23) Vinyl Acetate	5.85	43	1364	Below Cal	#	77
25) 2-Butanone	6.79	43	4453	0.468	ug/l #	83
31) Cyclohexane	7.63	56	29800	1.434	ug/l #	25
36) 1,1-Dichloropropene	7.63	75	89909	5.016	ug/l #	52
38) Carbon Tetrachloride	7.63	117	98394	5.458	ug/l #	17
43) Isopropyl Acetate	8.13	43	382937	10.774	ug/l	99
48) Methyl methacrylate	9.24	41	717998	39.578	ug/l #	53
49) 1,4-Dioxane	9.16	88	321	1.303	ug/l #	18
51) 4-Methyl-2-Pentanone	9.95	43	32342	1.219	ug/l	91
54) cis-1,3-Dichloropropene	9.87	75	122837	5.433	ug/l #	68
57) 1,3-Dichloropropane	10.73	76	24839	1.096	ug/l #	45
59) 2-Hexanone	10.74	43	442947	30.748	ug/l #	55
74) N-amyl acetate	12.01	43	14830	0.612	ug/l #	53
76) 1,2,3-Trichloropropane	12.38	75	455293	30.916	ug/l #	100
78) n-propylbenzene	12.58	91	1157	0.302	ug/l	91
80) 1,3,5-Trimethylbenzene	12.71	105	14072	0.371	ug/l	97
81) trans-1,4-Dichloro-2-buten	12.38	75	455293	88.411	ug/l #	4
93) 1,2,4-Trichlorobenzene	14.89	180	387	2.768	ug/l	82
94) Hexachlorobutadiene	14.98	225	75	Below Cal	#	45
95) Naphthalene	15.11	128	613	2.089	ug/l #	75


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070620\
Data File : VN062391.D
Acq On : 7 Jul 2020 2:48
Operator : JC/MD
Sample : PB129462TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129462TB

Quant Time: Jul 07 05:24:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 07 05:08:51 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1821

Delta R.T. 0.00 min

Lab File: VN062391.D

Acq: 7 Jul 2020 2:48

Instrument :

MSVOA_N

ClientSampleId :


PB129462TB

Tot Ion:168 Resp: 905746

Ion Ratio Lower Upper

168 100

99 64.7 52.9 79.3



Abundance Ion 167.90 (167.60 to 168.60): VN062365.D (-1799) (-)

Ion 98.90 (98.60 to 99.60): VN062365.D (-1799) (-)

7.63

400000



300000

200000

100000

0

Time--> 7.50 7.60 7.70



#8

Diethyl Ether

Concen: 0.317 ug/l

RT: 3.38 min Scan# 500

Delta R.T. 0.01 min

Lab File: VN062391.D


Acq: 7 Jul 2020 2:48

Tgt Ion: 74 Resp: 2200

Ion Ratio Lower Upper

74 100

45 107.8 59.9 179.7



Abundance Ion 74.00 (73.70 to 74.70): VN062391.D (-405) (-)

Ion 45.00 (44.70 to 45.70): VN062391.D (-405) (-)

3.38

3000

2500

2000


1500


1000

500

0

Time--> 3.36 3.37 3.38 3.39 3.40








#11

Tert butyl alcohol
 Concen: 0.557 ug/l
 RT: 4.75 min Scan# 926
 Delta R.T. 0.02 min
 Lab File: VN062391.D
 Acq: 7 Jul 2020 2:48

Instrument :
 MSVOA_N
 ClientSampled :
 PB129462TB




Tot Ion: 59 Resp: 1264
 Ion Ratio Lower Upper
 59 100
 57 14.6 8.5 12.7#




#13

Acrolein
 Concen: 0.466 ug/l
 RT: 3.56 min Scan# 556
 Delta R.T. -0.01 min
 Lab File: VN062391.D
 Acq: 7 Jul 2020 2:48

Tgt Ion: 56 Resp: 125
 Ion Ratio Lower Upper
 56 100
 55 0.0 59.6 89.4#





#16

Acetone

Concen: 8.188 ug/l

RT: 3.79 min Scan# 626

Delta R.T. 0.01 min

Lab File: VN062391.D

Acq: 7 Jul 2020 2:48

Instrument :

MSVOA_N

ClientSampleId :





PB129462TB

Tot Ion: 43 Resp: 61851

Ion Ratio Lower Upper

43 100

58 27.8 19.8 29.8



#18

Methyl Acetate

Concen: 0.729 ug/l

RT: 4.28 min Scan# 781

Delta R.T. 0.01 min

Lab File: VN062391.D




Acq: 7 Jul 2020 2:48


Tgt Ion: 43 Resp: 18967

Ion Ratio Lower Upper

43 100

74 16.2 14.2 21.4









#20
Methvlene Chloride
Concen: 12.457 ug/l
RT: 4.50 min Scan# 848
Delta R.T. -0.00 min
Lab File: VN062391.D
Acq: 7 Jul 2020 2:48

Instrument :
MSVOA_N
ClientSampleId :
PB129462TB

Tot Ion:	84	Resp:	149933
Ion	Ratio	Lower	Upper
84	100		
49	155.9	121.8	182.6
51	40.5	35.0	52.4
86	66.1	50.3	75.5




Abundance Ion 83.90 (83.60 to 84.60): VNO
Ion 48.90 (48.60 to 49.60): VNO
Ion 50.90 (50.60 to 51.60): VNO
Ion 85.90 (85.60 to 86.60): VNO






#23
Vinyl Acetate
Concen: Below Cal
RT: 5.85 min Scan# 1269
Delta R.T. 0.01 min
Lab File: VN062391.D
Acq: 7 Jul 2020 2:48

Tgt Ion:	43	Resp:	1364
Ion	Ratio	Lower	Upper
43	100		
86	0.0	6.6	9.8#



Abundance Ion 43.00 (42.70 to 43.70): VNO
Ion 86.00 (85.70 to 86.70): VNO









#25
2-Butanone
Concen: 0.468 ug/l
RT: 6.79 min Scan# 1560
Delta R.T. 0.00 min
Lab File: VN062391.D
Acq: 7 Jul 2020 2:48




Instrument :
MSVOA_N
ClientSampleId :
PB129462TB


Tot Ion: 43 Resp: 4453
Ion Ratio Lower Upper
43 100
72 12.7 16.4 24.6#



#31
Cyclohexane
Concen: 1.434 ug/l
RT: 7.63 min Scan# 1821
Delta R.T. 0.02 min
Lab File: VN062391.D
Acq: 7 Jul 2020 2:48

Tgt Ion: 56 Resp: 29800
Ion Ratio Lower Upper
56 100
69 149.4 23.7 35.5#
84 82.7 57.2 85.8





#33

1,2-Dichloroethane-d4

Concen: 50.212 ug/l

RT: 7.99 min Scan# 1933

Delta R.T. 0.00 min

Lab File: VN062391.D

Acq: 7 Jul 2020 2:48

Instrument :

MSVOA_N

ClientSampleId :


PB129462TB

Tot Ion: 65 Resp: 822672

Ion Ratio Lower Upper




65 100

67 51.9 0.0 101.0



Abundance Ion 64.90 (64.60 to 65.60): VN062391.D (-1839) (-)

Ion 66.90 (66.60 to 67.60): VN062391.D (-1839) (-)



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.55 min Scan# 2108

Delta R.T. -0.00 min

Lab File: VN062391.D

Acq: 7 Jul 2020 2:48


Tgt Ion: 114 Resp: 1705924

Ion Ratio Lower Upper

114 100

63 25.1 0.0 49.0



88 17.1 0.0 34.0




Abundance Ion 113.90 (113.60 to 114.60): VN062391.D (-2015) (-)

Ion 63.00 (62.70 to 63.70): VN062391.D (-2015) (-)

Ion 87.90 (87.60 to 88.60): VN062391.D (-2015) (-)









#35
 Dibromofluoromethane
 Concen: 48.682 ug/l
 RT: 7.55 min Scan# 1796
 Delta R.T. -0.00 min
 Lab File: VN062391.D
 Acq: 7 Jul 2020 2:48




Instrument :
 MSVOA_N
 ClientSampleId :
 PB129462TB


Tot Ion	Ratio	Lower	Upper
113	100		
111	103.4	83.2	124.8
192	17.1	14.2	21.2



#36
 1,1-Dichloropropene
 Concen: 5.016 ug/l
 RT: 7.63 min Scan# 1821
 Delta R.T. -0.12 min
 Lab File: VN062391.D
 Acq: 7 Jul 2020 2:48

Tgt Ion	Ratio	Lower	Upper
75	100		
110	8.7	15.7	47.1#
77	0.3	24.3	36.5#





#38

Carbon Tetrachloride

Concen: 5.458 ug/l

RT: 7.63 min Scan# 1821

Delta R.T. -0.11 min

Lab File: VN062391.D

Acq: 7 Jul 2020 2:48

Instrument :

MSVOA_N

ClientSampled :

PB129462TB


Tot Ion:117 Resp: 98394

Ion Ratio Lower Upper

117 100

119 5.3 75.9 113.9#




121 0.0 24.6 36.8#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 10.774 ug/l

RT: 8.13 min Scan# 1977

Delta R.T. 0.01 min

Lab File: VN062391.D

Acq: 7 Jul 2020 2:48


Tgt Ion: 43 Resp: 382937

Ion Ratio Lower Upper

43 100

61 15.9 13.1 19.7



87 8.8 7.5 11.3




Abundance Ion 43.00 (42.70 to 43.70): V

Ion 61.00 (60.70 to 61.70): V

Ion 87.00 (86.70 to 87.70): V









#48
Methvl methacrylate
Concen: 39.578 ug/l
RT: 9.24 min Scan# 2321
Delta R.T. 0.07 min
Lab File: VN062391.D
Acq: 7 Jul 2020 2:48




Instrument :
MSVOA_N
ClientSampled :
PB129462TB


Tot Ion: 41 Resp: 717998
Ion Ratio Lower Upper
41 100
69 0.1 51.5 77.3#
39 54.9 51.2 76.8



#49
1,4-Dioxane
Concen: 1.303 ug/l
RT: 9.16 min Scan# 2297
Delta R.T. -0.01 min
Lab File: VN062391.D
Acq: 7 Jul 2020 2:48

Tgt Ion: 88 Resp: 321
Ion Ratio Lower Upper
88 100
43 0.0 34.0 51.0#
58 0.0 63.0 94.4#





#50

Toluene-d8

Concen: 50.239 ug/l

RT: 10.06 min Scan# 2577

Delta R.T. -0.00 min

Lab File: VN062391.D

Acq: 7 Jul 2020 2:48

Instrument :

MSVOA_N

ClientSampled :





PB129462TB

Tot Ion: 98 Resp: 2185612

Ion Ratio Lower Upper

98 100

100 63.5 50.2 75.2



#51

4-Methyl-2-Pentanone

Concen: 1.219 ug/l

RT: 9.95 min Scan# 2544

Delta R.T. 0.00 min

Lab File: VN062391.D




Acq: 7 Jul 2020 2:48


Tgt Ion: 43 Resp: 32342

Ion Ratio Lower Upper

43 100

58 39.9 27.6 41.4





#54

cis-1,3-Dichloropropene

Concen: 5.433 ug/l

RT: 9.87 min Scan# 2519

Delta R.T. 0.06 min

Lab File: VN062391.D

Acq: 7 Jul 2020 2:48

Instrument :

MSVOA_N

ClientSampleId :

PB129462TB


Tot Ion: 75 Resp: 122837

Ion Ratio Lower Upper

75 100

77 0.0 26.0 39.0#




39 52.6 55.3 82.9#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 76.90 (76.60 to 77.60): VN0

Ion 39.00 (38.70 to 39.70): VN0



#57

1,3-Dichloropropane

Concen: 1.096 ug/l

RT: 10.73 min Scan# 2787

Delta R.T. 0.06 min

Lab File: VN062391.D


Acq: 7 Jul 2020 2:48

Tgt Ion: 76 Resp: 24839

Ion Ratio Lower Upper



76 100


78 0.6 25.0 37.4#



Abundance Ion 75.90 (75.60 to 76.60): VN0

Ion 77.90 (77.60 to 78.60): VN0









#59
2-Hexanone
Concen: 30.748 ug/l
RT: 10.74 min Scan# 2788
Delta R.T. 0.02 min
Lab File: VN062391.D
Acq: 7 Jul 2020 2:48




Instrument :
MSVOA_N
ClientSampleId :
PB129462TB


Tot Ion: 43 Resp: 442947
Ion Ratio Lower Upper
43 100
58 16.9 23.5 70.6#



#62
4-Bromofluorobenzene
Concen: 46.844 ug/l
RT: 12.38 min Scan# 3298
Delta R.T. 0.00 min
Lab File: VN062391.D
Acq: 7 Jul 2020 2:48

Tgt Ion: 95 Resp: 780488
Ion Ratio Lower Upper
95 100
174 73.3 0.0 145.8
176 68.8 0.0 140.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2988
Delta R.T. -0.00 min
Lab File: VN062391.D
Acq: 7 Jul 2020 2:48




Instrument :
MSVOA_N
ClientSampleId :
PB129462TB


Tot Ion:117 Resp: 1672532
Ion Ratio Lower Upper
117 100
82 61.7 49.4 74.2
119 31.8 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.32 min Scan# 3591
Delta R.T. -0.00 min
Lab File: VN062391.D
Acq: 7 Jul 2020 2:48

Tgt Ion:152 Resp: 609702
Ion Ratio Lower Upper
152 100
115 62.9 30.5 91.5
150 160.4 0.0 346.8





#74

N-amvl acetate

Concen: 0.612 ug/l

RT: 12.01 min Scan# 3184

Delta R.T. -0.03 min

Lab File: VN062391.D

Acq: 7 Jul 2020 2:48

Instrument :

MSVOA_N

ClientSampled :

PB129462TB

Tot Ion: 43 Resp: 14830


Ion Ratio Lower Upper

43 100

70 2.3 25.9 38.9#

55 2.4 17.4 26.0#

61 0.0 15.6 23.4#






Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 70.00 (69.70 to 70.70): VN0

Ion 55.00 (54.70 to 55.70): VN0

Ion 61.00 (60.70 to 61.70): VN0



#76

1,2,3-Trichloropropane

Concen: 30.916 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. -0.15 min

Lab File: VN062391.D


Acq: 7 Jul 2020 2:48

Tgt Ion: 75 Resp: 455293

Ion Ratio Lower Upper



75 100


77 1.0 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0





#78

n-propylbenzene

Concen: 0.302 ug/l

RT: 12.58 min Scan# 3360

Delta R.T. 0.01 min

Lab File: VN062391.D

Acq: 7 Jul 2020 2:48

Instrument :

MSVOA_N

ClientSampleId :





PB129462TB

Tot Ion: 91 Resp: 1157

Ion Ratio Lower Upper

91 100

120 26.1 10.9 32.6



#80

1,3,5-Trimethylbenzene

Concen: 0.371 ug/l

RT: 12.71 min Scan# 3400

Delta R.T. -0.00 min

Lab File: VN062391.D




Acq: 7 Jul 2020 2:48


Tgt Ion: 105 Resp: 14072

Ion Ratio Lower Upper

105 100

120 44.2 23.2 69.6





#81

trans-1,4-Dichloro-2-butene

Concen: 88.411 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. 0.10 min

Lab File: VN062391.D

Acq: 7 Jul 2020 2:48

Instrument :

MSVOA_N

ClientSampleId :

PB129462TB


Tot Ion: 75 Resp: 455293

Ion Ratio Lower Upper

75 100

53 0.3 92.3 138.5#




89 0.0 37.0 55.4#



Abundance Ion 74.90 (74.60 to 75.60): VN062391.D (-3173) (-)

Ion 53.00 (52.70 to 53.70): VN062391.D (-3173) (-)

Ion 88.90 (88.60 to 89.60): VN062391.D (-3173) (-)



#93

1,2,4-Trichlorobenzene

Concen: 2.768 ug/l

RT: 14.89 min Scan# 4080

Delta R.T. 0.01 min

Lab File: VN062391.D

Acq: 7 Jul 2020 2:48


Tgt Ion: 180 Resp: 387

Ion Ratio Lower Upper

180 100

182 71.8 47.6 142.8



145 31.3 15.9 47.7




Abundance Ion 179.80 (179.50 to 180.50): VN062391.D (-3984) (-)

Ion 181.80 (181.50 to 182.50): VN062391.D (-3984) (-)

Ion 144.80 (144.50 to 145.50): VN062391.D (-3984) (-)





#94

Hexachlorobutadiene

Concen: Below Cal

RT: 14.98 min Scan# 4106

Delta R.T. -0.01 min

Lab File: VN062391.D

Acq: 7 Jul 2020 2:48

Instrument :

MSVOA_N

ClientSampled :

PB129462TB


Tot Ion:225 Resp: 75

Ion Ratio Lower Upper

225 100

223 0.0 31.7 95.1#

227 40.0 30.3 90.8






Abundance

Ion 224.70 (224.40 to 225.40): V

Ion 222.70 (222.40 to 223.40): V

Ion 226.70 (226.40 to 227.40): V



#95

Naphthalene

Concen: 2.089 ug/l

RT: 15.11 min Scan# 4149

Delta R.T. 0.01 min

Lab File: VN062391.D

Acq: 7 Jul 2020 2:48


Tgt Ion:128 Resp: 613

Ion Ratio Lower Upper

128 100

127 4.7 10.8 16.2#

129 0.0 8.6 12.8#



Abundance

Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

