

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070620\
 Data File : VN062377.D
 Acq On : 6 Jul 2020 21:08
 Operator : JC/MD
 Sample : L3192-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-MW180-SOIL-604-608

Quant Time: Jul 07 06:45:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 07 05:08:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	958580	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	1723488	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	1649275	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	616590	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	861787	49.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.40%	
35) Dibromofluoromethane	7.55	113	591555	50.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.62%	
50) Toluene-d8	10.06	98	2224633	50.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.24%	
62) 4-Bromofluorobenzene	12.38	95	783512	46.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.10%	
Target Compounds						
16) Acetone	3.78	43	81018	10.135	ug/l	95
20) Methylene Chloride	4.50	84	104146	7.655	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070620\
Data File : VN062377.D
Acq On : 6 Jul 2020 21:08
Operator : JC/MD
Sample : L3192-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-MW180-SOIL-604-608

Quant Time: Jul 07 06:45:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 07 05:08:51 2020
Response via : Initial Calibration

