

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070620\
 Data File : VN062379.D
 Acq On : 6 Jul 2020 21:57
 Operator : JC/MD
 Sample : L3192-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-MW180-SOIL-614-616

Quant Time: Jul 07 06:40:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 07 05:08:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	970890	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	1742152	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	1672394	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	628347	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	874139	49.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.54%	
35) Dibromofluoromethane	7.55	113	584794	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
50) Toluene-d8	10.06	98	2248315	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
62) 4-Bromofluorobenzene	12.38	95	790712	46.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.94%	
Target Compounds						
16) Acetone	3.78	43	71487	8.829	ug/l	100
20) Methylene Chloride	4.50	84	165481	12.872	ug/l	97
79) 2-Chlorotoluene	12.65	91	67081	1.988	ug/l	96
82) 4-Chlorotoluene	12.75	91	61147	1.772	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070620\
Data File : VN062379.D
Acq On : 6 Jul 2020 21:57
Operator : JC/MD
Sample : L3192-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-MW180-SOIL-614-616

Quant Time: Jul 07 06:40:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 07 05:08:51 2020
Response via : Initial Calibration

