

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070620\
 Data File : VN062380.D
 Acq On : 6 Jul 2020 22:21
 Operator : JC/MD
 Sample : L3151-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-MW180-SOIL-584-588

Quant Time: Jul 07 06:37:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 07 05:08:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	945826	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	1690519	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	1633290	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	642243	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	846911	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
35) Dibromofluoromethane	7.55	113	572174	49.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.22%	
50) Toluene-d8	10.06	98	2173662	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
62) 4-Bromofluorobenzene	12.38	95	793474	48.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.12%	
Target Compounds						
20) Methylene Chloride	4.50	84	46855	2.665	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070620\
Data File : VN062380.D
Acq On : 6 Jul 2020 22:21
Operator : JC/MD
Sample : L3151-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-MW180-SOIL-584-588

Quant Time: Jul 07 06:37:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 07 05:08:51 2020
Response via : Initial Calibration

