

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070620\
 Data File : VN062382.D
 Acq On : 6 Jul 2020 23:09
 Operator : JC/MD
 Sample : L3205-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-214

Quant Time: Jul 07 06:32:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 07 05:08:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	940375	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	1687095	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	1696225	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	643734	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	862121	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
35) Dibromofluoromethane	7.55	113	573509	49.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.66%	
50) Toluene-d8	10.06	98	2234964	51.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.90%	
62) 4-Bromofluorobenzene	12.38	95	813139	49.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.70%	
Target Compounds						
16) Acetone	3.78	43	136826	17.447	ug/l	100
20) Methylene Chloride	4.51	84	135454	10.643	ug/l	97
43) Isopropyl Acetate	8.13	43	253804	7.220	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070620\
Data File : VN062382.D
Acq On : 6 Jul 2020 23:09
Operator : JC/MD
Sample : L3205-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-214

Quant Time: Jul 07 06:32:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 07 05:08:51 2020
Response via : Initial Calibration

