

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070620\
 Data File : VN062386.D
 Acq On : 7 Jul 2020 00:46
 Operator : JC/MD
 Sample : PB129462ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129462ZHE#02

Quant Time: Jul 07 05:49:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 07 05:08:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	928262	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	1686244	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	1666696	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	615710	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	841098	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
35) Dibromofluoromethane	7.55	113	572245	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
50) Toluene-d8	10.06	98	2228473	51.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.64%	
62) 4-Bromofluorobenzene	12.38	95	788316	47.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.74%	
Target Compounds						
16) Acetone	3.78	43	60347	7.796	ug/l	99
18) Methyl Acetate	4.29	43	25429	1.122	ug/l #	76
20) Methylene Chloride	4.50	84	150705	12.189	ug/l	96
43) Isopropyl Acetate	8.13	43	382216	10.879	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070620\
Data File : VN062386.D
Acq On : 7 Jul 2020 00:46
Operator : JC/MD
Sample : PB129462ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129462ZHE#02

Quant Time: Jul 07 05:49:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 07 05:08:51 2020
Response via : Initial Calibration

