

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070620\
 Data File : VN062387.D
 Acq On : 7 Jul 2020 1:11
 Operator : JC/MD
 Sample : PB129462ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129462ZHE#03

Quant Time: Jul 07 05:52:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 07 05:08:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	883192	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	1642013	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	1569304	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	581953	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	796354	49.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.70%	
35) Dibromofluoromethane	7.55	113	550081	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
50) Toluene-d8	10.06	98	2125456	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
62) 4-Bromofluorobenzene	12.38	95	752365	46.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.82%	
Target Compounds						
16) Acetone	3.78	43	69755	9.471	ug/l #	84
20) Methylene Chloride	4.50	84	138245	11.697	ug/l	95
43) Isopropyl Acetate	8.13	43	372856	10.898	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070620\
Data File : VN062387.D
Acq On : 7 Jul 2020 1:11
Operator : JC/MD
Sample : PB129462ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129462ZHE#03

Quant Time: Jul 07 05:52:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 07 05:08:51 2020
Response via : Initial Calibration

