

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070620\
 Data File : VN062389.D
 Acq On : 7 Jul 2020 1:59
 Operator : JC/MD
 Sample : PB129462ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129462ZHE#05

Quant Time: Jul 07 05:59:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 07 05:08:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	916365	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	1727122	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	1670153	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	641197	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	831902	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
35) Dibromofluoromethane	7.55	113	561285	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
50) Toluene-d8	10.06	98	2223555	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
62) 4-Bromofluorobenzene	12.38	95	799303	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
Target Compounds						
16) Acetone	3.78	43	57058	7.466	ug/l	95
18) Methyl Acetate	4.29	43	30851	1.503	ug/l #	76
20) Methylene Chloride	4.50	84	71111	5.034	ug/l	97
43) Isopropyl Acetate	8.13	43	388150	10.786	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070620\
Data File : VN062389.D
Acq On : 7 Jul 2020 1:59
Operator : JC/MD
Sample : PB129462ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129462ZHE#05

Quant Time: Jul 07 05:59:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 07 05:08:51 2020
Response via : Initial Calibration

