

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070620\
 Data File : VN062393.D
 Acq On : 7 Jul 2020 3:36
 Operator : JC/MD
 Sample : L3202-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RO-COMP

Quant Time: Jul 07 06:15:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 07 05:08:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	915560	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	1684965	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	1675264	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	613432	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	834115	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
35) Dibromofluoromethane	7.55	113	568861	49.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.98%	
50) Toluene-d8	10.06	98	2209056	51.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.82%	
62) 4-Bromofluorobenzene	12.38	95	796742	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
Target Compounds						
16) Acetone	3.78	43	75032	9.827	ug/l	92
18) Methyl Acetate	4.29	43	29276	1.401	ug/l	94
20) Methylene Chloride	4.51	84	127517	10.241	ug/l	96
43) Isopropyl Acetate	8.13	43	319602	9.104	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070620\
Data File : VN062393.D
Acq On : 7 Jul 2020 3:36
Operator : JC/MD
Sample : L3202-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RO-COMP

Quant Time: Jul 07 06:15:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 07 05:08:51 2020
Response via : Initial Calibration

