


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN070620\
 Data File : VN062361.D
 Acq On : 6 Jul 2020 11:48
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
 Title : SW846 8260
 Last Update : Tue Jul 07 05:08:51 2020



AutoFind: Scans 3297, 3298, 3299; Background Corrected with Scan 3281

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	26.9	53596	PASS
75	95	30	60	57.4	114503	PASS
95	95	100	100	100.0	199466	PASS
96	95	5	9	5.9	11726	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	72.3	144277	PASS
175	174	5	9	8.2	11896	PASS
176	174	95	101	96.6	139434	PASS
177	176	5	9	6.6	9269	PASS