


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN070620\
Data File : VN062369.D
Acq On : 6 Jul 2020 17:19
Operator : JC/MD
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070620W.M
Title : SW846 8260
Last Update : Tue Jul 07 05:08:51 2020



AutoFind: Scans 3297, 3298, 3299; Background Corrected with Scan 3282

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	25.3	50915	PASS
75	95	30	60	58.7	118197	PASS
95	95	100	100	100.0	201514	PASS
96	95	5	9	6.2	12582	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	73.9	148906	PASS
175	174	5	9	7.4	10958	PASS
176	174	95	101	95.2	141696	PASS
177	176	5	9	5.5	7755	PASS