

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070622\
 Data File : VN073333.D
 Acq On : 06 Jul 2022 09:46
 Operator : JC\MD
 Sample : VN0706WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0706WBL01

Quant Time: Jul 07 04:20:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	237697	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	416449	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	354997	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	143404	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	178059	51.891	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.780%
35) Dibromofluoromethane	7.951	113	140759	50.803	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.600%
50) Toluene-d8	10.381	98	461173	46.864	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.720%
62) 4-Bromofluorobenzene	12.669	95	169350	44.782	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070622\
Data File : VN073333.D
Acq On : 06 Jul 2022 09:46
Operator : JC\MD
Sample : VN0706WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0706WBL01

Quant Time: Jul 07 04:20:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

