

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070622\
 Data File : VN073346.D
 Acq On : 06 Jul 2022 15:36
 Operator : JC\MD
 Sample : N3564-16
 Misc : 4.87g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-16-2-6

Quant Time: Jul 07 04:23:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.010	168	280471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	467038	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	466562	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	252725	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	169804	41.938	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.880%
35) Dibromofluoromethane	7.945	113	139837	45.003	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.000%
50) Toluene-d8	10.374	98	518088	46.945	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.900%
62) 4-Bromofluorobenzene	12.668	95	247211	58.290	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.580%
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.769	43	106052	16.651	ug/l #	89
25) 2-Butanone	7.275	43	28526	7.884	ug/l #	85
31) Cyclohexane	7.998	56	10880	1.641	ug/l #	67
39) Methylcyclohexane	9.386	83	35918	6.034	ug/l	96
52) Toluene	10.439	92	380150	41.688	ug/l	100
68) m/p-Xylenes	11.892	106	20406	2.823	ug/l	85
84) 1,2,4-Trimethylbenzene	13.310	105	36405	2.050	ug/l	90
85) sec-Butylbenzene	13.439	105	33537	1.566	ug/l #	57
89) n-Butylbenzene	13.880	91	35061	2.432	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070622\
Data File : VN073346.D
Acq On : 06 Jul 2022 15:36
Operator : JC\MD
Sample : N3564-16
Misc : 4.87g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-16-2-6

Quant Time: Jul 07 04:23:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

