

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070622\
 Data File : VN073349.D
 Acq On : 06 Jul 2022 16:49
 Operator : JC\MD
 Sample : N3586-02 10X
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31268

Quant Time: Jul 07 04:24:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	340950	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	561314	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	541704	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	267382	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	206102	41.874	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	83.740%		
35) Dibromofluoromethane	7.951	113	173791	46.537	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.080%		
50) Toluene-d8	10.380	98	596569	44.978	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	89.960%		
62) 4-Bromofluorobenzene	12.669	95	278369	54.612	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	109.220%		
Target Compounds						
52) Toluene	10.445	92	20038	1.828	ug/l	98
84) 1,2,4-Trimethylbenzene	13.304	105	35026	1.864	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070622\
Data File : VN073349.D
Acq On : 06 Jul 2022 16:49
Operator : JC\MD
Sample : N3586-02 10X
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31268

Quant Time: Jul 07 04:24:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

