

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070622\
 Data File : VN073350.D
 Acq On : 06 Jul 2022 17:14
 Operator : JC\MD
 Sample : N3586-03 10X
 Misc : 1.35g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31473

Quant Time: Jul 07 04:24:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	333163	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	556563	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	536249	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	265553	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	202514	42.106	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.220%
35) Dibromofluoromethane	7.951	113	172398	46.558	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.120%
50) Toluene-d8	10.380	98	590510	44.901	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.800%
62) 4-Bromofluorobenzene	12.668	95	273828	54.180	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.360%
Target Compounds						
					Qvalue	
31) Cyclohexane	7.992	56	30125	3.824	ug/l #	56
52) Toluene	10.439	92	739048	68.010	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070622\
Data File : VN073350.D
Acq On : 06 Jul 2022 17:14
Operator : JC\MD
Sample : N3586-03 10X
Misc : 1.35g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31473

Quant Time: Jul 07 04:24:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

