

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070721\
 Data File : VN067647.D
 Acq On : 7 Jul 2021 20:48
 Operator : JC/MD
 Sample : M2773-19
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-7

Manual Integrations
 APPROVED

MMDadoda
 7/8/2021 6:04:51 PM

Quant Time: Jul 08 06:10:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	324091	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	595729	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	573281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.680	152	217652	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	254673	51.995	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 104.000%			
35) Dibromofluoromethane	8.024	113	177107	49.207	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 98.420%			
50) Toluene-d8	10.446	98	780119	51.782	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.560%			
62) 4-Bromofluorobenzene	12.733	95	254077	47.741	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 95.480%			
Target Compounds						
11) Tert butyl alcohol	5.385	59	7727m	10.978	ug/l	Qvalue
16) Acetone	4.299	43	50972	33.073	ug/l	97
20) Methylene Chloride	5.095	84	41335	10.945	ug/l	94
25) 2-Butanone	7.343	43	36771	14.368	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070721\
Data File : VN067647.D
Acq On : 7 Jul 2021 20:48
Operator : JC/MD
Sample : M2773-19
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-7

Quant Time: Jul 08 06:10:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

MMDadoda
7/8/2021 6:04:51 PM

