

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070721\
 Data File : VN067648.D
 Acq On : 7 Jul 2021 21:12
 Operator : JC/MD
 Sample : PB137507TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB137507TB

Manual Integrations
 APPROVED

MMDadoda
 7/8/2021 6:04:52 PM

Quant Time: Jul 08 06:10:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	311083	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	586510	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	653454	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	240553	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	249254	53.017	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 106.040%			
35) Dibromofluoromethane	8.024	113	173208	48.880	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 97.760%			
50) Toluene-d8	10.443	98	813059	54.816	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 109.640%			
62) 4-Bromofluorobenzene	12.734	95	287728	54.914	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 109.820%			
Target Compounds					Qvalue	
16) Acetone	4.296	43	55639	37.611	ug/l	98
20) Methylene Chloride	5.093	84	25559m	6.800	ug/l	
43) Isopropyl Acetate	8.566	43	46577	5.016	ug/l #	92
95) Naphthalene	15.512	128	4900	3.403	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070721\
Data File : VN067648.D
Acq On : 7 Jul 2021 21:12
Operator : JC/MD
Sample : PB137507TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB137507TB

Manual Integrations
APPROVED

MMDadoda
7/8/2021 6:04:52 PM

Quant Time: Jul 08 06:10:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

