

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070722\
 Data File : VN073361.D
 Acq On : 07 Jul 2022 10:07
 Operator : JC\MD
 Sample : VN0707WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0707WBL01

Quant Time: Jul 22 05:33:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	310804	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	515635	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	485137	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	233401	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	195934	43.669	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.340%
35) Dibromofluoromethane	7.951	113	164829	48.047	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.100%
50) Toluene-d8	10.380	98	542895	44.557	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.120%
62) 4-Bromofluorobenzene	12.668	95	242961	51.888	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070722\
Data File : VN073361.D
Acq On : 07 Jul 2022 10:07
Operator : JC\MD
Sample : VN0707WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0707WBL01

Quant Time: Jul 22 05:33:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

