

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070819\
 Data File : VN056608.D
 Acq On : 8 Jul 2019 14:45
 Operator : JC/SP
 Sample : PB121169TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB121169TB

Quant Time: Jul 09 04:10:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	297310	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	509128	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	528160	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	212099	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	171688	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
35) Dibromofluoromethane	7.59	113	152118	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
50) Toluene-d8	10.09	98	642001	51.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.90%	
62) 4-Bromofluorobenzene	12.40	95	239052	58.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.04%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.79	59	6740	12.145	ug/l #	90
16) Acetone	3.82	43	49565	30.023	ug/l	100
20) Methylene Chloride	4.55	84	8297	1.808	ug/l	100
25) 2-Butanone	6.84	43	12436	5.692	ug/l	99
29) Tetrahydrofuran	7.21	42	56763	42.511	ug/l	99
43) Isopropyl Acetate	8.17	43	21825	3.214	ug/l #	89
95) Naphthalene	15.13	128	5724	5.773	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070819\
Data File : VN056608.D
Acq On : 8 Jul 2019 14:45
Operator : JC/SP
Sample : PB121169TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB121169TB

Quant Time: Jul 09 04:10:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

