

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070820\
 Data File : VN062417.D
 Acq On : 8 Jul 2020 19:27
 Operator : JC/MD
 Sample : L3215-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-17

Quant Time: Jul 09 02:55:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 09 02:12:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	178657	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	316316	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	305015	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	108926	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	140890	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
35) Dibromofluoromethane	7.55	113	104578	53.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.98%	
50) Toluene-d8	10.06	98	413997	53.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.28%	
62) 4-Bromofluorobenzene	12.38	95	146036	51.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.54%	
Target Compounds						
30) Chloroform	7.33	83	4440	1.117	ug/l	95
64) Tetrachloroethene	10.61	164	4577	2.529	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070820\
Data File : VN062417.D
Acq On : 8 Jul 2020 19:27
Operator : JC/MD
Sample : L3215-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-17

Quant Time: Jul 09 02:55:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 09 02:12:34 2020
Response via : Initial Calibration

