

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070820\
 Data File : VN062431.D
 Acq On : 9 Jul 2020 1:59
 Operator : JC/MD
 Sample : L3215-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-8

Quant Time: Jul 09 03:29:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 09 02:12:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	169316	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	305011	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	294094	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	101160	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	134591	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
35) Dibromofluoromethane	7.55	113	98666	52.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.66%	
50) Toluene-d8	10.06	98	393599	52.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.76%	
62) 4-Bromofluorobenzene	12.38	95	138439	50.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.80%	
Target Compounds						
16) Acetone	3.78	43	2504	2.265	ug/l	92
30) Chloroform	7.33	83	23126	6.138	ug/l	96
64) Tetrachloroethene	10.61	164	2708	1.552	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070820\
Data File : VN062431.D
Acq On : 9 Jul 2020 1:59
Operator : JC/MD
Sample : L3215-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-8

Quant Time: Jul 09 03:29:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 09 02:12:34 2020
Response via : Initial Calibration

