

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070820\
 Data File : VN062432.D
 Acq On : 9 Jul 2020 2:23
 Operator : JC/MD
 Sample : L3215-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-9

Quant Time: Jul 09 04:14:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 09 02:12:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	174842	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	311992	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	301957	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	106529	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	139514	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
35) Dibromofluoromethane	7.55	113	101564	52.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.32%	
50) Toluene-d8	10.06	98	405627	53.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.56%	
62) 4-Bromofluorobenzene	12.38	95	143871	51.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.42%	
Target Compounds						
16) Acetone	3.79	43	4520	3.959	ug/l	98
44) Trichloroethene	8.80	130	682	0.332	ug/l	92
64) Tetrachloroethene	10.60	164	6417	3.581	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070820\
Data File : VN062432.D
Acq On : 9 Jul 2020 2:23
Operator : JC/MD
Sample : L3215-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-9

Quant Time: Jul 09 04:14:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 09 02:12:34 2020
Response via : Initial Calibration

