

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070820\
 Data File : VN062434.D
 Acq On : 9 Jul 2020 3:11
 Operator : JC/MD
 Sample : L3215-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1A

Quant Time: Jul 09 04:21:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 09 02:12:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	183780	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	331335	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	311449	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	117206	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	143253	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
35) Dibromofluoromethane	7.55	113	106350	51.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.86%	
50) Toluene-d8	10.06	98	428602	53.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.02%	
62) 4-Bromofluorobenzene	12.38	95	151904	51.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.82%	
Target Compounds						
					Qvalue	
16) Acetone	3.78	43	43010	35.838	ug/l	100
25) 2-Butanone	6.80	43	7105	4.054	ug/l	93
39) Methylcyclohexane	9.05	83	9900	2.778	ug/l	99
67) Ethyl Benzene	11.49	91	6878	0.605	ug/l #	78
68) m/p-Xylenes	11.60	106	3816	0.949	ug/l	93
69) o-Xylene	11.92	106	1301	0.332	ug/l	92
84) 1,2,4-Trimethylbenzene	13.02	105	3725	0.504	ug/l	90
95) Naphthalene	15.11	128	12608	2.407	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070820\
Data File : VN062434.D
Acq On : 9 Jul 2020 3:11
Operator : JC/MD
Sample : L3215-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1A

Quant Time: Jul 09 04:21:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 09 02:12:34 2020
Response via : Initial Calibration

