

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070820\
 Data File : VN062437.D
 Acq On : 9 Jul 2020 4:23
 Operator : JC/MD
 Sample : L3164-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-MW180-GW-588-590

Quant Time: Jul 09 05:20:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 09 02:12:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	183457	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	331293	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	315804	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	112811	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	144439	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery =	102.38%		
35) Dibromofluoromethane	7.55	113	106007	51.77	ug/l	0.00
Spiked Amount 50.000			Recovery =	103.54%		
50) Toluene-d8	10.06	98	429966	53.19	ug/l	0.00
Spiked Amount 50.000			Recovery =	106.38%		
62) 4-Bromofluorobenzene	12.38	95	149514	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery =	101.22%		

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.83	85	1165	0.623	ug/l	94
9) 1,1,2-Trichlorotrifluoroet	3.73	101	9591	5.186	ug/l	97
12) 1,1-Dichloroethene	3.70	96	6111	3.245	ug/l	84
16) Acetone	3.79	43	3764	3.142	ug/l	90
27) cis-1,2-Dichloroethene	6.79	96	6213	2.629	ug/l	82
30) Chloroform	7.33	83	9112	2.232	ug/l	96
44) Trichloroethene	8.80	130	802896	367.869	ug/l	98
55) 1,1,2-Trichloroethane	10.54	97	3915	1.752	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070820\
Data File : VN062437.D
Acq On : 9 Jul 2020 4:23
Operator : JC/MD
Sample : L3164-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-MW180-GW-588-590

Quant Time: Jul 09 05:20:20 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 09 02:12:34 2020
Response via : Initial Calibration

