

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN070821\  
 Data File : VN067654.D  
 Acq On : 8 Jul 2021 11:05  
 Operator : JC/MD  
 Sample : VN0708MBL01  
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_N/MEOH  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0708MBL01

Quant Time: Jul 08 15:50:20 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N070721W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jul 07 16:18:53 2021  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	8.088	168	381805	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	705514	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	658164	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	233632	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	295341	51.183	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	102.360%	
35) Dibromofluoromethane	8.024	113	212318	49.810	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	99.620%	
50) Toluene-d8	10.443	98	905933	50.775	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	101.560%	
62) 4-Bromofluorobenzene	12.734	95	294177	46.674	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	93.340%	
Target Compounds						
95) Naphthalene	15.509	128	32363	5.671	ug/l	Qvalue 99
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN070821\  
 Data File : VN067654.D  
 Acq On : 8 Jul 2021 11:05  
 Operator : JC/MD  
 Sample : VN0708MBL01  
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_N/MEOH  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0708MBL01

Quant Time: Jul 08 15:50:20 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N070721W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jul 07 16:18:53 2021  
 Response via : Initial Calibration

