

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070821\
 Data File : VN067663.D
 Acq On : 8 Jul 2021 15:03
 Operator : JC/MD
 Sample : M2773-19
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-7

Manual Integrations
 APPROVED

MMDadoda
 7/9/2021 5:43:12 PM

Quant Time: Jul 09 03:53:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	336004	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	628140	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	601446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	230484	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	265794	52.342	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 104.680%			
35) Dibromofluoromethane	8.024	113	187444	49.391	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 98.780%			
50) Toluene-d8	10.443	98	820012	51.621	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.240%			
62) 4-Bromofluorobenzene	12.733	95	268160	47.787	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 95.580%			
Target Compounds						Qvalue
11) Tert butyl alcohol	5.409	59	7271m	9.964	ug/l	
16) Acetone	4.296	43	55692	34.854	ug/l	96
20) Methylene Chloride	5.095	84	43234	11.048	ug/l	91
25) 2-Butanone	7.348	43	37043	13.961	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070821\
Data File : VN067663.D
Acq On : 8 Jul 2021 15:03
Operator : JC/MD
Sample : M2773-19
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-7

Manual Integrations
APPROVED

MMDadoda
7/9/2021 5:43:12 PM

Quant Time: Jul 09 03:53:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

