

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070821\
 Data File : VN067664.D
 Acq On : 8 Jul 2021 15:28
 Operator : JC/MD
 Sample : M2799-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-3

Quant Time: Jul 09 03:53:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	323852	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	602968	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	575666	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	218570	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	258084	52.730	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.460%
35) Dibromofluoromethane	8.024	113	181096	49.711	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.420%
50) Toluene-d8	10.446	98	789803	51.795	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.600%
62) 4-Bromofluorobenzene	12.734	95	255609	47.452	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.900%
Target Compounds						
						Qvalue
16) Acetone	4.299	43	15987	10.381	ug/l	99
20) Methylene Chloride	5.095	84	17161	4.135	ug/l	89
95) Naphthalene	15.507	128	3648	3.333	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070821\
Data File : VN067664.D
Acq On : 8 Jul 2021 15:28
Operator : JC/MD
Sample : M2799-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-3

Quant Time: Jul 09 03:53:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

