

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070821\
 Data File : VN067666.D
 Acq On : 8 Jul 2021 16:17
 Operator : JC/MD
 Sample : PB137565TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB137565TB

Quant Time: Jul 09 03:54:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	314316	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	591108	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	643948	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	231937	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	247585	52.120	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 104.240%		
35) Dibromofluoromethane	8.024	113	175109	49.032	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 98.060%		
50) Toluene-d8	10.446	98	802976	53.715	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 107.440%		
62) 4-Bromofluorobenzene	12.733	95	281271	53.264	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 106.520%		
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	37391	25.015	ug/l	97
20) Methylene Chloride	5.101	84	31369	8.411	ug/l	91
43) Isopropyl Acetate	8.566	43	44131	4.716	ug/l #	89
95) Naphthalene	15.509	128	19257	4.606	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070821\
Data File : VN067666.D
Acq On : 8 Jul 2021 16:17
Operator : JC/MD
Sample : PB137565TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB137565TB

Quant Time: Jul 09 03:54:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

