

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070821\
 Data File : VN067671.D
 Acq On : 8 Jul 2021 18:21
 Operator : JC/MD
 Sample : M2967-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1-8FT-STEFANIC-AVE-TP-2

Quant Time: Jul 09 03:56:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	343255	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	657384	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	696288	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	249025	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	269573	51.965	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 103.920%		
35) Dibromofluoromethane	8.024	113	189451	47.700	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 95.400%		
50) Toluene-d8	10.443	98	885994	53.294	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 106.580%		
62) 4-Bromofluorobenzene	12.733	95	301606	51.356	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 102.720%		
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	72892	44.655	ug/l	99
18) Methyl Acetate	4.867	43	12075	2.634	ug/l	91
20) Methylene Chloride	5.100	84	31126	7.578	ug/l	89
43) Isopropyl Acetate	8.563	43	48526	4.662	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070821\
Data File : VN067671.D
Acq On : 8 Jul 2021 18:21
Operator : JC/MD
Sample : M2967-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1-8FT-STEFANIC-AVE-TP-2

Quant Time: Jul 09 03:56:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

