

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070821\
 Data File : VN067672.D
 Acq On : 8 Jul 2021 18:45
 Operator : JC/MD
 Sample : M2967-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1-5FT-CADMUS-AVE-TP-3

Quant Time: Jul 09 03:56:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	348322	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	660803	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	705621	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	255061	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	273671	51.987	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.980%	
35) Dibromofluoromethane	8.024	113	193530	48.474	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 96.940%	
50) Toluene-d8	10.443	98	886774	53.065	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.120%	
62) 4-Bromofluorobenzene	12.733	95	303319	51.381	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 102.760%	
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	66736	40.289	ug/l	99
18) Methyl Acetate	4.865	43	16542	3.555	ug/l	97
20) Methylene Chloride	5.098	84	30303	7.241	ug/l	90
43) Isopropyl Acetate	8.563	43	48642	4.649	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070821\
Data File : VN067672.D
Acq On : 8 Jul 2021 18:45
Operator : JC/MD
Sample : M2967-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1-5FT-CADMUS-AVE-TP-3

Quant Time: Jul 09 03:56:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

