

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070821\
 Data File : VN067674.D
 Acq On : 8 Jul 2021 19:34
 Operator : JC/MD
 Sample : M2968-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Manual Integrations
 APPROVED

MMDadoda
 7/9/2021 5:43:06 PM

Quant Time: Jul 09 03:57:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	334558	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	631954	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	580759	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	206808	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	262537	51.924	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 103.840%			
35) Dibromofluoromethane	8.024	113	188753	49.436	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 98.880%			
50) Toluene-d8	10.443	98	811828	50.797	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 101.600%			
62) 4-Bromofluorobenzene	12.733	95	249521	44.197	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 88.400%			
Target Compounds						Qvalue
18) Methyl Acetate	4.859	43	13150	2.943	ug/l #	87
20) Methylene Chloride	5.095	84	4277m	0.462	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070821\
Data File : VN067674.D
Acq On : 8 Jul 2021 19:34
Operator : JC/MD
Sample : M2968-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Manual Integrations
APPROVED

MMDadoda
7/9/2021 5:43:06 PM

Quant Time: Jul 09 03:57:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

