

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070822\
 Data File : VN073389.D
 Acq On : 08 Jul 2022 13:39
 Operator : JC\MD
 Sample : N3586-01DL 5X
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31267DL

Quant Time: Jul 08 13:58:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.010	168	299755	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	496353	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	469331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	237166	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	184848	42.717	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.440%
35) Dibromofluoromethane	7.945	113	154984	46.932	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.860%
50) Toluene-d8	10.380	98	527734	44.995	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.000%
62) 4-Bromofluorobenzene	12.669	95	247101	54.823	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.640%
Target Compounds						
					Qvalue	
52) Toluene	10.439	92	172103	17.759	ug/l	98
84) 1,2,4-Trimethylbenzene	13.304	105	61477	3.689	ug/l	96
89) n-Butylbenzene	13.880	91	15442	1.142	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070822\
Data File : VN073389.D
Acq On : 08 Jul 2022 13:39
Operator : JC\MD
Sample : N3586-01DL 5X
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31267DL

Quant Time: Jul 08 13:58:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

