

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070822\
 Data File : VN073403.D
 Acq On : 08 Jul 2022 19:22
 Operator : JC\MD
 Sample : N3642-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Jul 09 05:30:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	296348	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	503957	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	475112	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	225648	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	194554	45.477	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.960%
35) Dibromofluoromethane	7.951	113	161470	48.158	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.320%
50) Toluene-d8	10.380	98	533793	44.825	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.660%
62) 4-Bromofluorobenzene	12.669	95	235907	51.549	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070822\
Data File : VN073403.D
Acq On : 08 Jul 2022 19:22
Operator : JC\MD
Sample : N3642-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Jul 09 05:30:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

